


# Schedule of Accreditation

issued by

## United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

 <b>UKAS</b> CALIBRATION  <b>27974</b>  Accredited to ISO/IEC 17025:2017	<b>Positive Weighing Solutions Ltd</b>	
	Issue No: 001    Issue date: 11 July 2023	
	<b>Unit 8</b> The Clifton Centre Spring Lane South Malvern Worcestershire WR14 1BJ	<b>Contact: Mrs Lucy Bennett-Poole</b> Tel: +44 (0) 1684 561002 E-Mail: <a href="mailto:info@positiveweighing.co.uk">info@positiveweighing.co.uk</a> Website: <a href="http://www.positiveweighing.co.uk">www.positiveweighing.co.uk</a>
Calibration performed by the Organisation at the locations specified		

### Locations covered by the organisation and their relevant activities

#### Laboratory locations:

Location details	Activity	Location code
<b>Address</b> Unit 8 The Clifton Centre Spring Lane South Malvern Worcestershire WR14 1BJ	<b>Local contact</b> Mrs Lucy Bennett-Poole Tel: +44 (0)1684 561002 Email: <a href="mailto:info@positiveweighing.co.uk">info@positiveweighing.co.uk</a> Website: <a href="http://www.positiveweighing.co.uk">www.positiveweighing.co.uk</a>	Mass  P

#### Site activities performed away from the locations listed above:

Location details	Activity	Location code
The customers' site or premises must be suitable for the nature of the particular calibrations undertaken and will be the subject of contract review arrangements between the laboratory and the customer.	Mass	S



27974

Accredited to  
ISO/IEC 17025:2017

**Schedule of Accreditation**  
issued by  
**United Kingdom Accreditation Service**  
2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

**Positive Weighing Solutions Ltd**

**Issue No: 001 Issue date: 11 July 2023**

**Calibration performed by the Organisation at the locations specified**

Calibration and Measurement Capability (CMC)

Measured Quantity Instrument or Gauge	Range	Expanded Measurement Uncertainty ( $k = 2$ )	Remarks	Location Code
<b>MASS</b>				
Non-Automatic Weighing Machines See notes 1 and 2	200 mg 500 mg 1 g 2 g 5 g 10 g 20 g 50 g 100 g 200 g 500 g 1 kg 2 kg 5 kg 10 kg 20 kg 50 kg 100 kg 200 kg 500 kg 1000 kg 1500 kg	0.0092 mg 0.012 mg 0.015 mg 0.018 mg 0.024 mg 0.033 mg 0.045 mg 0.074 mg 0.14 mg 0.28 mg 0.69 mg 1.4 mg 3.8 mg 9.6 mg 20 mg 40 mg 130 mg 1.6 g 16 g 39 g 78 g 120 g	Methods Consistent with EURAMET CG18  Note 1. Weights are available in OIML Class:  E2 1 mg to 1 kg. Max grouped load 2.0 kg  F1 1 mg to 20 kg. Max grouped load 40 kg  M1 500 mg to 20 kg. Max grouped load 1500 kg.  Note 2. Other loads within the overall listed range may also be used.	P & S
<b>END</b>				



27974

Accredited to  
ISO/IEC 17025:2017

## Schedule of Accreditation

issued by

### United Kingdom Accreditation Service

2 Pine Trees, Chertsey Lane, Staines-upon-Thames, TW18 3HR, UK

### Positive Weighing Solutions Ltd

Issue No: 001 Issue date: 11 July 2023

Calibration performed by the Organisation at the locations specified

## Appendix - Calibration and Measurement Capabilities

### Introduction

The definitive statement of the accreditation status of a calibration laboratory is the Accreditation Certificate and the associated Schedule of Accreditation. This Schedule of Accreditation is a critical document, as it defines the measurement capabilities, ranges and boundaries of the calibration activities for which the organisation holds accreditation.

### Calibration and Measurement Capabilities (CMCs)

The capabilities provided by accredited calibration laboratories are described by the Calibration and Measurement Capability (CMC), which expresses the lowest measurement uncertainty that can be achieved during a calibration. If a particular device under calibration itself contributes significantly to the uncertainty (for example, if it has limited resolution or exhibits significant non-repeatability) then the uncertainty quoted on a calibration certificate will be increased to account for such factors.

The CMC is normally used to describe the uncertainty that appears in an accredited calibration laboratory's schedule of accreditation and is the uncertainty for which the laboratory has been accredited using the procedure that was the subject of assessment. The measurement uncertainty is calculated according to the procedures given in the GUM and is normally stated as an expanded uncertainty at a coverage probability of 95 %, which usually requires the use of a coverage factor of  $k = 2$ . An accredited laboratory is not permitted to quote an uncertainty that is smaller than the published measurement uncertainty in certificates issued under its accreditation.

### Expression of CMCs - symbols and units

It should be noted that the percentage symbol (%) represents the number 0.01. In cases where the measurement uncertainty is stated as a percentage, this is to be interpreted as meaning percentage of the measurand. Thus, for example, a measurement uncertainty of 1.5 % means  $1.5 \times 0.01 \times q$ , where  $q$  is the quantity value.

The notation  $Q[a, b]$  stands for the root-sum-square of the terms between brackets:  $Q[a, b] = [a^2 + b^2]^{1/2}$